


Flow Diagram of SIMION Trajectory Calculations with Locations of User Program Segment Calls



(a) These are repeated for each active array instance (highest priority instance containing the current ion): first the active electrostatic array (if any), then the active magnetic array (if any). That is 0, 1, or 2 times.

(i) These are repeated for each ion currently being processed. That is one ion or (if Grouped Mode) all ions.